

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsaco.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

REF: KEMSA/PROC/OIT 5/16-17

3rd April 2017

To all Bidders,

CLARIFICATION NO. 1

Dear Sir/Madam,

RE: TENDER NO. KEMSA/OIT 5/2016-2018 FOR SUPPLY OF NON PHARMACEUTICALS (SYRINGES).

The following are queries raised by interested bidder and responses made as below:

Question: Kindly please clarify: What does RUP mean. Also Specifications mentions: Syringes should not have a rubber gasket for the following syringes;

- Item 1 - Syringes 2PC- 10ML RUPwith G 21 needle X 1½"
- Item 2 - Syringes 2PC- 2ML RUP with G23 needle X 1"
- Item 3 - Syringes 2PC - 5ML RUP with G 21 needle X 1½"
- Item 4 Syringes (without needle) 60mls, with luer slide (feeding)
- Item 5 - Syringes 2PC- 20ML with G21 needle X 1½"

Response: RUP means Re - Use Prevention. Item 1, 2, 3 and 5 should completely not have a Rubber Gasket since all the syringes are two piece.

For item 4 - Syringes (without needle) 60mls, with luer slide, the rubber gasket should be latex free

Yours faithfully,


Charles E. Juma

Director Procurement

For: Ag. Chief Executive Officer