

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsa.co.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

REF: KEMSA/PROC/ONT/2018

14th May 2018

To all Bidders,

CLARIFICATION NO. 2

Dear Sir/Madam,

**RE: TENDER NO. KEMSA/ONT04/2017-2019- SUPPLY OF NON-PHARMACEUTICALS
(Surgical sutures, Dispensing envelopes)**

The following are queries raised by an interested bidder and responses made as below:-

Question 1: For item no. 1 and 2. Please clarify if the NEEDLE LENGTH is 60mm and 90mm respectively

Answer: The specifications for item No.1 and 2 remain as they are NEEDLE LENGTH is 60mm and 90mm respectively

Question 2: For item no 20. A) Thickness (GSM) of Paper used in Envelope.
B) Sealing flap adhesive type.

- Dry Adhesive **OR**
- Peel and Seal Adhesive.

Answer: A) The (GSM) thickness of the paper used in envelope should be 0.22mm
B) The sealing flap adhesive should be peel and seal adhesive

Yours faithfully,

Charles E. Juma
Director Procurement
For: Ag. Chief Executive Officer