

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsaco.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA/GF-PROC 001/18 (92)

Date: 8th August, 2018

ADDENDUM 2

To: All Prospective Bidders

Dear Sir/Madam,

RE: TENDER NO.: GF ATM- HIV NFM -18/19- ONT- 01 - SUPPLY OF BLOOD SAFETY COMMODITIES

In accordance with ITT Clause 11.1 (Clarification of Tender Documents) of the Open National Tender (ONT) issued under the above reference, we hereby respond to queries raised by prospective bidders.

Question 1.

1. Vacuum Tube – Vacutainers are blood collection tubes where a sample is collected with a vacuum BCT needle.
2. Non – Vacuum tube – Vacutainers where blood is collected from a syringe and transferred to the tube.

Kindly clarify whether Item No. 2, 3 and 4 are Vacuum Tubes vacutainers or Non-Vacuum Tubes Vacutainers

Response:

The Vacutainer should be a vacuum vacutainer as blood is drawn by vacuum from the donor blood bag needle.

Question 2:

In regard to addendum 1 of the above named tender, for **sample submission**, do you need the unique Identifiers and Barcodes to be put on the primary and secondary package label or they can be put thereafter if we get the award.

Response:

Unique identifiers and Barcodes to be put thereafter in the event of an award.

Yours faithfully,

A handwritten signature in blue ink, consisting of several overlapping, scribbled lines that form a stylized, illegible name.

Edward Buluma
FOR: CHIEF EXECUTIVE OFFICER