

# KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area  
P.O. Box 47715, 00100 GPO,  
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400  
GSM +254 719 033000, +254 733 606600  
Email: info@kemsaco.ke

*All Correspondence should be addressed to Chief Executive Officer*

When replying please quote our ref:

REF: KEMSA/PROC/OIT6/2018

25<sup>th</sup> October 2018

To all Bidders,

## ADDENDUM NO. 1

Dear Sir/Madam,

### RE: TENDER NO. KEMSA/OIT 06/2018-2020 SUPPLY OF X-RAY CONSUMABLES

The following Query was asked by an interested bidder and answer was given as detailed below.

#### **Question:**

For item No. 5 (Thermal Printing Paper 110mm X 18 metres) - Please note that all brands that are 110mm X 18 meters can only give a minimum of 200 prints. On the specification it's stated on the packaging parameters that each roll to provided a minimum of 240 prints. Please clarify on this point.

#### **Answer:**

Thermal printing paper should range between 110mm X 18m - 110mm x 20m and minimum of 200 prints - 270 prints.

Yours faithfully,

A handwritten signature in blue ink, appearing to read 'Caroline Gichinga', is written over a light blue horizontal line.

**Caroline Gichinga**  
**FOR: CHIEF EXECUTIVE OFFICER**