

# KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area  
P.O. Box 47715, 00100 GPO,  
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400  
GSM +254 719 033000, +254 733 606600  
Email: info@kemsa.co.ke

*All Correspondence should be addressed to Chief Executive Officer*

When replying please quote our ref:

REF: KEMSA/PROC/OIT 04/2018-2020

Date: October 8, 2018

## ADDENDUM NO 1

To all Bidders

Dear Sir/Madam,

### RE: ADDENDUM TO TENDER NO. KEMSA/OIT-04/2018-2020 FOR SUPPLY OF LABORATORY TEST KITS

Below are the specifications for **Item 2 - Syphilis Anti-body Screening Test** in the subject tender which had earlier been omitted.

ITEM DESCRIPTION	SPECIFICATION
Syphilis Anti-body Screening Test	<p><b>Product Parameters</b></p> <ul style="list-style-type: none"><li>• Should be bases on flocculation principle using non treponemal antigens</li><li>• Qualitative determination of reagin antibodies in serum or plasma for sero-diagnosis of syphilis.</li><li>• Must be WHO prequalified</li><li>• Should be calibrated to WHO reference serum</li><li>• Should be suitable to use with either serum or plasma.</li><li>• Sensitivity should be &gt; 70%</li><li>• Should be rapid preferably &lt; 15 min</li><li>• Should have positive and negative controls</li><li>• The kit should have all accessories required for the test such as Cards, Dropper, Applicator</li><li>• The kit should have a shelf life of minimal 18-24 months at the time of delivery.</li><li>• Preferable Pack size of 50 to 200 tests.</li><li>• Storage temperature for the kit should be +2 to +8oC</li></ul> <p><b>Packaging Parameters:</b></p> <ul style="list-style-type: none"><li>• Should be a complete kit of 50 – 100 tests</li><li>• Standard weight of carton should be ≤ 20kg.</li></ul>

ITEM DESCRIPTION	SPECIFICATION
	<p><b>Labeling Parameters:</b></p> <ul style="list-style-type: none"> <li>• Labeling should be in English in indelible ink.</li> <li>• Generic/ chemical name clearly indicated</li> <li>• Quantity/volume matches the label name</li> <li>• Labeled "KEMSA "using indelible ink upon award of contract</li> <li>• Each carton to be clearly marked with the name and characteristics of the article and number of units per carton.</li> <li>• Manufacturer's Name and address, Country of Origin, Batch No, Date of Manufacture and expiry.</li> <li>• The products should have a minimum of 75% remaining shelf-life at the time of delivery.</li> <li>• Should conform to KEBS / ISO standard OR equivalent.</li> <li>• Manufacturer must be KEBS / ISO certified or equivalent</li> </ul> <p>Submission of sample:  <b>Submit a complete kit of 100 tests for evaluation</b></p>

Yours faithfully,



**CAROLINE GICHINGA**  
**FOR: CHIEF EXECUTIVE OFFICER**