

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsaco.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA/GF-PROC/2019 (072)

Date: 29th March, 2019

ADDENDUM 1:

Attn.: All Prospective Bidders

RE: TENDER: GF ATM HIV NFM 18/19 - OIT 013 SUPPLY OF MALE CONDOMS (II)

In accordance with ITT Clause 11.1 (Clarification of Tender Documents) of the Open International Tender (OIT) issued under the above reference, we hereby respond to questions raised by prospective bidders as follows:

Question 1:

The tender document mentions that the foil size must be 56mm. x 56mm. Our standard foil size is 58x58. Please advise if this is acceptable.

Response:

Kindly adhere to the technical specifications.

Question 2:

The tender document mentions the condom with form 49 and 53mm. Pls kindly confirm the quantity allocated of each size.

Response:

There is no split of quantities. Condoms Width: 53mm with a tolerance of ± 2 mm and Length: 180 minimum are required.

Question 3:

As we have never received the artwork from KEMSA before, pls kindly confirm if the samples can be submitted in plain foil and plain gross box.

Response:

Artwork will be provided by NASCOP after award

All other terms and conditions of the tender remain the same.

Yours Faithfully



John Kabuchi

FOR : CHIEF EXECUTIVE OFFICER